
Compressed Oxygen Cylinder, certified – Specifications

Product	: Compressed Oxygen Cylinder, certified
Standard	: Source bulk manufactured in compliance with EU GMP.
Applicable Monograph	: Batch tested to the specifications depicted below, taken from EP 0417.
Clarifying Notes	: Non-Medicinal; Finished product outside scope of Certificate of GMP Compliance of a Manufacturer and Manufacturing Licence (GMP).
Expiry (Gas, Cylinders)	: 36 months

Parameter	Limit
Purity	≥ 99.50 %
Moisture Content	NMT 67 ppm
Carbon Monoxide	NMT 5 ppm
Carbon Dioxide	NMT 300 ppm